

EC-DECLARATION OF CONFORMITY

Manufacturer: Address:	Hitachi, Ltd. 2-16-1, Higashi-Ueno, Taito-ku, Tokyo, 110-0015, Japan		
selected conformity assessment procedure:	Annex II excludin	g (4) RoHS Article 7 ((b), Module A
EU Authorized representative: Address:	Hitachi Medical Systems GmbH Otto-von-Guericke-Ring 3, D-65205 Wiesbaden, Germany		
Product:	Diagnostic Ultrasound System		
Model Code:	Prosound α6		
Classification (MDD, Annex IX): Classification rule (MDD, Annex IX) Statement:		ategories (RoHS(II), Annex I):	No.8

We are exclusively responsible for the declaration of conformity and herewith declare that the above mentioned products including all its options meet the provisions of the following EC Council Directives and Standards. All supporting documentations are retained under the premises of the manufacturer.

DIRECTIVES

General applicable directives:

Medical Device Directive : Council Directive 93/42/EEC as amended by 2007/47/EC; MDD).

Standards : MDD Harmonized Standards (published in the Office Journal of the European Communities) applicable to this product are :

EN 60601-1:2006/AC:2010, EN 60601-1-2:2007/AC:2010, EN 60601-2-37:2008 EN 62366:2008, EN 62304:2006:/AC:2008, EN 60601-1-6:2010, EN 1041:2008 EN ISO 14971:2012

Other Standards : IEC 62079:2001, ISO 7010:2003/A6:2010, ISO 3864-2:2004, ISO 15223-1:2012 IEC 60601-1-9:2007

Notified body :TÜV SÜD Product Service GmbH is Notified Body with identification no. 0123Address (for MDD):Ridlerstraße 65, 80339 München, Germany

RoHS Directive : Directive 2011/65/EU of 8 June 2011 concerning on the restriction of the use of certain hazardous substances in electrical and electronic equipment.

Standards : RoHS Directive Harmonized Standards (published in the Office Journal of the European Communities) applicable to this product is

EN 50581:2012

Hitachi Healthcare Manufacturing, Ltd. Tokyo Works 3-7-19 Imai, Ome-shi, Tokyo, 198-8577 Japan

Starting of CE Marking: 20581199

Jan. 04. 2017

Date: Signature:

Address:

Production facility :

naoda

Name of issuer : Ryosuke Maeda Position : Senior Manager Quality Assurance Department 1 Place: Tokyo, JAPAN