

## **EC-DECLARATION OF CONFORMITY**

Manufacturer: Address:	Hitachi, Ltd. 2-16-1 Higashi-Ueno, Taito-ku, Tokyo, 110-0015, Japan
Selected conformity assessment procedure: Annex II excluding (4) RoHS Article 7 (b), Module A	
EU Authorized representa Address:	tive: Hitachi Medical Systems GmbH Otto-von-Guericke-Ring 3, D-65205 Wiesbaden, Germany
Product: Model Code REF :	<b>Electronic Sector Probe</b> UST-52105
Classification (MDD, Annex IX): IIa Categories (RoHS(II), Annex I): No.8 Classification rule (MDD, Annex IX): Rule 10	
Statement: We are exclusively responsible for the declaration of conformity and herewith declare that the above-mentioned product including all its options meet the provisions of the following EC Council Directives and Standards. All supporting documentations are retained under the premises of the manufacturer.	
DIRECTIVES	
General applicable directives: Medical Device Directive : Council Directive 93/42/EEC of 14 June 1993 as amended by 2007/47/EC;	
	ÜV SÜD Product Service GmbH is Notified Body with identification no. 0123 idlerstraße 65, 80339 München, Germany
	irective 2011/65/EU of 8 June 2011 concerning on the restriction of the se of certain hazardous substances in electrical and electronic equipment.
	litachi Healthcare Manufacturing, Ltd.    Analytical Systems Kashiwa Factory -1, Shintoyofuta, Kashiwa-shi, Chiba, 277-0804 JAPAN

Starting of CE Marking: 206G8176 Date: Sep.03,2019

Signature:

Name of issuer : Position :

maeda Ryosuke Maeda

Management Representative

Place: Tokyo, JAPAN