

## EU-DECLARATION OF CONFORMITY

Manufacturer: FUJIFILM Healthcare Corporation  
Address: 2-1, Shintoyofuta, Kashiwa-shi, Chiba, 277-0804 Japan  
Actor ID/SRN: JP-MF-000018708

EU Authorized representative: **FUJIFILM Healthcare Deutschland GmbH**  
Address: **Otto-von-Guericke-Ring 3, D-65205 Wiesbaden, Germany**  
Actor ID/SRN: DE-AR-000017504

### Statement:

We are exclusively responsible for the declaration of conformity and herewith declare that the below mentioned products including all its options meet the provisions of the following EU Regulation and EC Council Directives and common specifications. All supporting documentations are retained under the premises of the manufacturer.

Basic UDI-DI; 457359621US\_SystemsWE  
referred to in Part C of Annex VI


Product: **Diagnostic Ultrasound System**

Model Code: **ARIETTA 650 (\*Include attachment sheet)**

Serial Number: G3171472

UMDNS Code:15-976 GMDN Code: 40761 EMDN Code: Z110401

Photograph for market plate:



Intended purpose :

It is intended to be used by qualified medical doctors and any qualified persons by national law for performing tomographic and hemodynamics diagnoses in the following parts of the human body:

Abdominal/ Cardiac/ Obstetrics/ Gynecology/ Superficial organs/ Urology/ Intraoperative

Classification/rule (MDR, Annex XIII): II a/Rule 10 Categories (RoHS( II ), Annex I): No.8

General applicable Regulation / Directives:

**Medical Device Regulation : (EU)/2017/745**

**RoHS Directive : Directive 2011/65/EU of 8 June 2011 and (EU) 2015/863 of 31 March 2015 concerning on the restriction of the use of certain hazardous substances in electrical and electronic equipment.**

References CS; Non Common Specification, because it has not been issued yet.

Standards;

MDR Harmonized Standards (published in the Official Journal of the European Union) applicable to this product are :  
No Harmonized Standards related to ARIETTA 650

Other Standards :

EN 60601-1:2006/A1:2013, EN 60601-1-2:2015, EN 62304:2006  
EN 60601-1-6:2010/A1:2015, EN 60601-2-37:2008/A1:2015, EN 62366:2008/A1:2015,  
EN ISO 14971:2012, EN ISO 10993-1:2020, EN 1041:2008/A1:2013,  
EN ISO 15223-1:2016, EN ISO 13485:2016/AC:2018

Standards : RoHS Directive Harmonized Standards (published in the Official Journal of the European Union) applicable to this product is :

EN IEC 63000:2018

Notified body : TÜV Rheinland LGA Products GmbH is Notified Body with identification no. 0197  
Address : Tillystraße 2, 90431 Nürnberg, Germany

Selected conformity assessment procedure:  
(for MDR) MDR : Annex IX

Certification Identification Number: HZ 2251444-2

Selected conformity assessment procedure  
(for RoHS) Article 7 (b), Module A




Additional Information;

Production facility : **FUJIFILM Healthcare Manufacturing Corporation**  
**Analytical Systems Kashiwa Factory**  
Address: **2-1, Shintoyofuta, Kashiwa-shi, Chiba, 277-0804 JAPAN**

Place: **Chiba, JAPAN**

Date: **JUN. 20, 2022**

Signature:   
Name of issuer : Shinichiro Kishi  
Position : Senior Manager  
Product Design Department

  
Shinichi Chiba  
Person responsible  
for regulatory compliance

Attachment sheet for Declaration of Conformity of ARIETTA 650.

This Declaration of Conformity is also effective to following marketing name(s)<sup>1</sup>.

**ARIETTA 650 DeepInsight**

<sup>1</sup> "marketing name is NOT printed in the product label and we use it to our product specification and brochure depend on system configuration.